
Midmark M11 - v1.0.5
Total Cycles: 355

Name: RITTER2
Sterilizer ID: V2313228
09/13/2021 10:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 181.3 F 3.4
4:00 189.7 F 7.8
6:00 208.3 F 13.1
8:00 221.8 F 7.7
10:00 234.8 F 13.3
12:00 239.5 F 11.1
14:00 238.9 F 10.2
16:00 252.9 F 17.1
18:00 264.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.4 F 29.3
2:00 271.5 F 29.0
3:00 272.7 F 29.3
4:00 271.3 F 28.7

Min 271.3 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 214.6 F 0.0
10:00 205.9 F 0.0
15:00 205.8 F 0.0
20:00 204.8 F 0.0
25:00 203.3 F 0.0
30:00 205.5 F 0.0

Drying Complete

FILLING: 1:37
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:34
CYCLE COMPLETE

09/13/2021 11:33 AM
ACCEPTED: X
REJECTED:
Operator: U01