
Midmark M11 - v1.0.0
Total Cycles: 1116

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2021 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.7 F 0.0
2:00 139.9 F 0.6
4:00 172.5 F 3.0
6:00 184.9 F 6.7
8:00 205.8 F 11.4
10:00 218.1 F 4.7
12:00 228.7 F 9.3
14:00 237.9 F 10.4
16:00 240.3 F 11.0
18:00 236.2 F 8.8
20:00 251.0 F 15.5
22:00 262.5 F 22.1
24:00 272.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 271.8 F 28.3
2:00 272.4 F 28.8
3:00 271.3 F 28.4
4:00 271.7 F 28.4

Min 271.1 F 28.1
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.2 F 0.0
5:00 217.9 F 0.0
10:00 199.6 F 0.0
15:00 196.7 F 0.0
20:00 192.9 F 0.0
Dry Cycle Aborted

FILLING: 1:02
HEATING: 24:18
STERILIZING: 4:00
VENTING: 3:04
DRYING: 22:09
TOTAL CYCLE: 00:54:32

08/25/2021 09:29 AM
ACCEPTED: X
REJECTED:
Operator: U02