
Midmark M11 - v1.0.5
Total Cycles: 153

Name: RITTER2
Sterilizer ID: V2313228
06/29/2021 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 174.6 F 2.8
4:00 189.4 F 7.8
6:00 213.6 F 9.3
8:00 225.4 F 10.1
10:00 231.7 F 7.7
12:00 240.7 F 11.0
14:00 246.7 F 14.1
16:00 259.7 F 21.2
18:00 271.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.4 F 29.5
2:00 272.3 F 29.9
3:00 271.5 F 29.1
4:00 272.1 F 29.4

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 232.7 F 0.0
10:00 226.8 F 0.0
15:00 232.5 F 0.0
20:00 228.1 F 0.0
25:00 232.0 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:30
STERILIZING: 4:00
VENTING: 2:49
DRYING: 30:00
TOTAL CYCLE: 00:57:20
CYCLE COMPLETE

06/29/2021 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____