
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000558
15:42 28/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 4:01
STERILIZING 5:17
136.0 C 229kPa 5:17
136.1 C 230kPa 5:47
136.1 C 230kPa 6:17
136.3 C 231kPa 6:47
136.4 C 231kPa 7:17
136.3 C 231kPa 7:47
136.1 C 230kPa 8:17
136.4 C 230kPa 8:47
136.4 C 230kPa 8:47
VENTING 8:47
AIR DRYING 9:41
CYCLE COMPLETE 24:41

Digital Signature #
7373BDCFBD13D61D
