
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008006
10:39 30/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:10
STERILIZING 4:21
136.4 C 223kPa 4:21
136.2 C 219kPa 4:51
136.4 C 223kPa 5:21
136.4 C 220kPa 5:51
136.5 C 223kPa 6:21
136.4 C 222kPa 6:51
136.1 C 220kPa 7:21
136.1 C 222kPa 7:51
136.1 C 222kPa 7:51
VENTING 7:51
AIR DRYING 8:46
CYCLE COMPLETE 26:52

Digital Signature #
C231A19E379013F8
