
Midmark M11 - v1.0.0
Total Cycles: 2455

Name: RITTER 1
Sterilizer ID: V2178198
10/19/2022 08:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.4 F 0.0
2:00 200.5 F 3.0
4:00 212.2 F 7.5
6:00 209.3 F 12.4
8:00 221.4 F 6.5
10:00 234.2 F 11.7
12:00 235.6 F 8.7
14:00 241.0 F 10.4
16:00 246.1 F 13.2
18:00 258.2 F 19.7
20:00 269.1 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 29.4
2:00 272.3 F 28.8
3:00 271.5 F 28.6
4:00 272.4 F 29.4

Min 271.4 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 230.0 F 0.0
10:00 235.2 F 0.0
15:00 234.1 F 0.0
20:00 226.8 F 0.0
25:00 225.5 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:51
CYCLE COMPLETE

10/19/2022 09:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____