
Midmark M11 - v1.0.0
Total Cycles: 1328

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.0 F 0.0
2:00 92.1 F 0.0
4:00 159.8 F 0.0
6:00 144.8 F 0.0
8:00 157.1 F 0.1
10:00 173.8 F 0.7
12:00 191.3 F 2.7
14:00 214.1 F 4.9
16:00 220.6 F 7.5
18:00 229.5 F 10.6
20:00 235.7 F 8.9
22:00 239.4 F 10.8
24:00 233.8 F 7.7
26:00 244.3 F 12.0
28:00 245.7 F 12.7
30:00 257.2 F 18.8
32:00 267.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.4 F 28.6
2:00 271.7 F 27.9
3:00 271.8 F 28.4
4:00 272.2 F 28.4

Min 271.2 F 27.7
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 194.2 F 0.0
10:00 205.2 F 0.0
15:00 204.8 F 0.0
20:00 203.7 F 0.0
25:00 204.5 F 0.0
30:00 204.6 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:29
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:11:41
CYCLE COMPLETE

11/02/2021 07:56 AM
ACCEPTED: X
REJECTED:
Operator: U01