
Midmark M11 - v1.0.0
Total Cycles: 4425

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2024 03:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.2 F 0.0
2:00 97.5 F 0.7
4:00 122.5 F 2.7
6:00 155.1 F 5.5
8:00 196.2 F 8.8
10:00 198.6 F 12.7
12:00 214.7 F 4.2
14:00 226.8 F 7.5
16:00 237.0 F 11.3
18:00 228.9 F 6.7
20:00 242.5 F 12.0
22:00 238.7 F 9.7
24:00 250.8 F 15.6
26:00 261.4 F 21.6
28:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.5 F 28.9
2:00 271.1 F 28.1
3:00 272.4 F 29.0
4:00 272.1 F 28.5

Min 271.0 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 209.8 F 0.0
10:00 192.9 F 0.0
15:00 195.0 F 0.0
20:00 218.3 F 0.0
25:00 237.9 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:41
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:07:13
CYCLE COMPLETE

04/03/2024 04:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____