

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001530  
12:24 10/25/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:01  
PRESSURIZING 6:46  
STERILIZING 8:08  
136.0 C 220kPa 8:08  
136.2 C 221kPa 8:38  
136.4 C 218kPa 9:08  
136.5 C 220kPa 9:38  
136.4 C 223kPa 10:08  
136.3 C 222kPa 10:38  
136.4 C 222kPa 11:09  
136.4 C 223kPa 11:38  
VENTING 11:38  
AIR DRYING 12:43  
CYCLE COMPLETE 42:44  
-----

Digital Signature #  
741DA716D3012DAC  
-----