

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3103

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2023 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.2 F 0.0  
2:00 70.5 F 0.0  
4:00 78.4 F 0.3  
6:00 112.2 F 1.8  
8:00 147.9 F 5.0  
10:00 179.1 F 8.9  
12:00 191.4 F 13.5  
14:00 220.0 F 6.3  
16:00 224.4 F 10.5  
18:00 229.8 F 7.5  
20:00 238.7 F 11.2  
22:00 235.0 F 8.7  
24:00 248.7 F 15.1  
26:00 260.7 F 21.8  
28:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.0  
1:00 272.8 F 30.1  
2:00 272.7 F 30.1  
3:00 272.7 F 30.1  
4:00 272.4 F 30.1

Min 270.8 F 28.8  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.9 F 0.0  
5:00 229.9 F 0.0  
10:00 224.8 F 0.0  
15:00 183.1 F 0.0  
20:00 178.7 F 0.0  
25:00 177.2 F 0.0  
30:00 175.9 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 28:30  
STERILIZING: 4:00  
VENTING: 3:02  
DRYING: 30:00  
TOTAL CYCLE: 01:06:20  
CYCLE COMPLETE

11/09/2023 10:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_