

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2404

Name: RITTER 1  
Sterilizer ID: V2178198  
10/05/2022 02:24 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.1 F 0.0  
2:00 188.3 F 3.1  
4:00 195.2 F 6.9  
6:00 220.5 F 11.7  
8:00 226.5 F 6.6  
10:00 236.9 F 11.6  
12:00 232.6 F 8.4  
14:00 240.4 F 10.9  
16:00 245.1 F 12.8  
18:00 257.0 F 19.2  
20:00 267.9 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.6 F 29.5  
2:00 271.9 F 28.7  
3:00 271.9 F 29.1  
4:00 272.5 F 29.3

Min 271.4 F 28.5  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.4 F 0.0  
5:00 231.8 F 0.0  
10:00 246.0 F 0.0  
15:00 245.1 F 0.0  
20:00 235.8 F 0.0  
25:00 225.1 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 21:24  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

10/05/2022 03:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_