
Midmark M11 - v1.0.0
Total Cycles: 1490

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.6 F 0.0
2:00 198.1 F 3.6
4:00 207.8 F 8.7
6:00 224.1 F 8.7
8:00 232.4 F 9.9
10:00 234.2 F 8.1
12:00 244.1 F 12.5
14:00 244.9 F 12.8
16:00 257.7 F 19.7
18:00 269.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.8
2:00 272.5 F 29.3
3:00 271.6 F 28.5
4:00 272.5 F 29.1

Min 271.6 F 28.5
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 220.9 F 0.0
10:00 233.3 F 0.0
15:00 235.4 F 0.0
20:00 229.5 F 0.0
25:00 227.3 F 0.0
30:00 225.4 F 0.0

Drying Complete

FILLING: 2:27
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

12/29/2021 04:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____