
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000479
14:53 26/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:17
STERILIZING 4:23
136.0 C 228kPa 4:23
136.3 C 231kPa 4:53
136.4 C 233kPa 5:23
136.2 C 231kPa 5:53
136.4 C 233kPa 6:23
136.3 C 232kPa 6:53
136.4 C 233kPa 7:23
136.4 C 231kPa 7:53
136.4 C 231kPa 7:53
VENTING 7:53
AIR DRYING 8:49
CYCLE COMPLETE 23:49

Digital Signature #
E4A9003449D77CFB
