
Midmark M11 - v1.0.5
Total Cycles: 2236

Name: RITTER2
Sterilizer ID: V2313228
04/12/2023 07:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.6 F 0.0
2:00 88.3 F 0.0
4:00 152.8 F 0.0
6:00 111.5 F 1.9
8:00 149.8 F 5.3
10:00 193.5 F 9.2
12:00 205.8 F 13.3
14:00 217.5 F 7.2
16:00 227.8 F 11.4
18:00 231.1 F 7.3
20:00 241.8 F 12.7
22:00 239.7 F 10.5
24:00 252.0 F 16.7
26:00 262.8 F 23.1
28:00 272.8 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 272.5 F 29.6
1:00 271.1 F 29.6
2:00 271.7 F 30.0
3:00 272.4 F 29.9
4:00 272.8 F 29.8

Min 271.0 F 29.5
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.6 F 0.0
5:00 222.7 F 0.0
10:00 219.0 F 0.0
15:00 220.8 F 0.0
20:00 222.8 F 0.0
25:00 234.6 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 28:45
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:06:43
CYCLE COMPLETE

04/12/2023 08:16 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____