
Midmark M11 - v1.0.5
Total Cycles: 2318

Name: RITTER2
Sterilizer ID: V2313228
05/01/2023 11:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.8 F 0.8
2:00 215.6 F 9.1
4:00 232.5 F 13.5
6:00 248.9 F 6.8

C573: HEATUP MODE
STEAM TEMP OVERLIM
CHAMBER TEMP 285F
CHAMBER PRESS 11.5PSI
** ITEMS NOT STERILE **

05/01/2023 11:26 AM