
Midmark M11 - v1.0.0
Total Cycles: 1182

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2021 01:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.6 F 0.0
2:00 148.0 F 2.0
4:00 173.3 F 5.4
6:00 202.1 F 10.0
8:00 222.4 F 4.8
10:00 234.7 F 9.2
12:00 238.8 F 9.5
14:00 241.9 F 11.5
16:00 239.0 F 9.6
18:00 252.5 F 16.5
20:00 264.3 F 23.5
22:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.9
2:00 273.1 F 29.3
3:00 272.6 F 29.0
4:00 272.6 F 28.7

Min 271.2 F 28.0
Max 273.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 220.7 F 0.0
10:00 214.1 F 0.0
15:00 200.6 F 0.0
20:00 187.5 F 0.0
25:00 191.4 F 0.0
30:00 193.0 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:03
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

09/16/2021 02:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____