
Midmark M11 - v1.0.5
Total Cycles: 3265

Name: RITTER2
Sterilizer ID: V2313228
12/29/2023 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 71.1 F | 0.0 |
| 2:00 | 72.9 F | 0.0 |
| 4:00 | 83.1 F | 0.2 |
| 6:00 | 118.0 F | 1.8 |
| 8:00 | 155.8 F | 5.5 |
| 10:00 | 182.3 F | 9.7 |
| 12:00 | 207.0 F | 6.5 |
| 14:00 | 219.2 F | 7.4 |
| 16:00 | 231.9 F | 11.9 |
| 18:00 | 232.9 F | 8.7 |
| 20:00 | 241.6 F | 11.2 |
| 22:00 | 245.7 F | 13.4 |
| 24:00 | 258.5 F | 20.3 |
| 26:00 | 270.0 F | 27.7 |

STERILIZING

| mm:ss | Degrees | PSI |
|-------|---------|------|
| 0:00 | 271.5 F | 28.8 |
| 1:00 | 271.8 F | 29.2 |
| 2:00 | 272.4 F | 29.6 |
| 3:00 | 272.5 F | 29.4 |
| 4:00 | 272.4 F | 29.2 |

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING

Drying Start

| mm:ss | Degrees | PSI |
|-------|---------|-----|
| 0:00 | 235.0 F | 0.0 |

Dry Cycle Aborted

FILLING: 0:58
HEATING: 26:45
STERILIZING: 4:00
VENTING: 2:46
DRYING: 1:52
TOTAL CYCLE: 00:36:20

12/29/2023 11:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____