

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5483

Name: RITTER 1  
Sterilizer ID: V2178198  
04/07/2025 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 113.3 F 0.0  
2:00 117.1 F 0.9  
4:00 123.6 F 3.4  
6:00 146.9 F 6.3  
8:00 175.1 F 9.8  
10:00 192.3 F 13.5  
12:00 208.6 F 4.1  
14:00 219.4 F 7.8  
16:00 230.9 F 11.8  
18:00 230.4 F 7.4  
20:00 241.7 F 12.3  
22:00 238.8 F 10.2  
24:00 250.5 F 15.8  
26:00 260.9 F 21.7  
28:00 270.6 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.4 F 29.4  
2:00 272.4 F 29.0  
3:00 271.4 F 28.6  
4:00 271.8 F 29.2

Min 271.3 F 28.6  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
5:00 192.2 F 0.0  
10:00 177.1 F 0.0  
15:00 186.3 F 0.0  
20:00 195.4 F 0.0  
25:00 203.5 F 0.0  
30:00 207.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:54  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:07:46  
CYCLE COMPLETE

04/07/2025 11:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_