
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000833
15:48 22/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:14
STERILIZING 4:45
136.4 C 233kPa 4:45
136.3 C 232kPa 5:15
136.3 C 232kPa 5:45
136.4 C 233kPa 6:15
136.4 C 233kPa 6:45
136.2 C 231kPa 7:15
136.4 C 231kPa 7:45
136.2 C 231kPa 8:15
136.2 C 231kPa 8:15
VENTING 8:15
AIR DRYING 9:12
CYCLE COMPLETE 14:42

Digital Signature #
939863D3B6B51F63
