
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000033
14:49 24/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 3:54
CYCLE FAULT 15
NOT STERILE
END TIME 3:54

Digital Signature #
53C28D1F145FCED5
