
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008127
14:04 02/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:07
STERILIZING 4:08
136.0 C 220kPa 4:08
136.4 C 225kPa 4:38
136.4 C 222kPa 5:08
136.2 C 221kPa 5:38
136.4 C 220kPa 6:08
136.4 C 223kPa 6:38
136.2 C 221kPa 7:08
136.2 C 222kPa 7:38
136.2 C 222kPa 7:38
VENTING 7:38
AIR DRYING 8:29
CYCLE COMPLETE 8:41

Digital Signature #
2D3A39579B1E060A
