
Midmark M11 - v1.0.5
Total Cycles: 2402

Name: RITTER2
Sterilizer ID: V2313228
05/23/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.7 F 0.0
2:00 170.0 F 3.5
4:00 182.3 F 7.8
6:00 203.7 F 12.8
8:00 221.0 F 7.0
10:00 233.9 F 12.3
12:00 235.5 F 9.8
14:00 236.5 F 9.0
16:00 248.6 F 15.0
18:00 260.3 F 21.7
20:00 271.3 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.7
2:00 271.6 F 28.9
3:00 272.6 F 29.8
4:00 271.9 F 29.0

Min 271.5 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 208.1 F 0.0
10:00 224.5 F 0.0
15:00 219.2 F 0.0
20:00 216.4 F 0.0
25:00 215.1 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:40
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

05/23/2023 03:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____