
Midmark M11 - v1.0.0
Total Cycles: 2753

Name: RITTER 1
Sterilizer ID: V2178198
01/09/2023 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.1 F 0.0
2:00 192.5 F 0.0
4:00 209.1 F 5.4
6:00 223.1 F 11.1
8:00 229.0 F 7.5
10:00 242.8 F 13.2
12:00 242.9 F 12.2
14:00 242.5 F 11.6
16:00 256.8 F 19.0
18:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.6 F 29.2
4:00 272.2 F 28.7

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.7 F 0.0
5:00 232.8 F 0.0
10:00 243.2 F 0.0
15:00 239.9 F 0.0
20:00 229.4 F 0.0
25:00 226.2 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 19:05
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:59
CYCLE COMPLETE

01/09/2023 04:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____