
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000853
16:12 07/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:20
STERILIZING 4:27
136.0 C 225kPa 4:27
136.2 C 232kPa 4:57
136.4 C 233kPa 5:27
136.4 C 231kPa 5:57
136.0 C 231kPa 6:27
136.2 C 232kPa 6:57
136.2 C 231kPa 7:27
136.3 C 232kPa 7:57
136.3 C 232kPa 7:57
VENTING 7:57
AIR DRYING 8:52
CYCLE COMPLETE 23:52

Digital Signature #
6825F2D852506DF9
