
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007503
8:14 06/06/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:40
STERILIZING 3:38
136.0 C 220kPa 3:38
136.4 C 225kPa 4:08
136.3 C 224kPa 4:38
136.4 C 224kPa 5:08
135.9 C 220kPa 5:38
136.4 C 224kPa 6:08
136.1 C 220kPa 6:38
136.4 C 223kPa 7:08
136.4 C 223kPa 7:08
VENTING 7:08
AIR DRYING 7:59
CYCLE COMPLETE 37:59

Digital Signature #
68C75B30F83A2B05
