
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 005016
15:18 30/01/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:02
STERILIZING 4:02
136.4 C 223kPa 4:02
136.4 C 224kPa 4:32
136.2 C 218kPa 5:02
136.4 C 223kPa 5:32
136.2 C 223kPa 6:02
136.2 C 225kPa 6:32
136.4 C 224kPa 7:02
136.2 C 224kPa 7:32
136.2 C 224kPa 7:32
VENTING 7:32
AIR DRYING 8:23
CYCLE COMPLETE 18:32

Digital Signature #
A3FA8E1483B3F235
