
Midmark M11 - v1.0.0
Total Cycles: 3275

Name: RITTER 1
Sterilizer ID: V2178198
05/24/2023 11:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.1 F 0.0
2:00 202.2 F 1.6
4:00 213.2 F 7.1
6:00 225.9 F 12.8
8:00 232.7 F 8.6
10:00 239.3 F 10.5
12:00 243.4 F 12.6
14:00 242.7 F 11.6
16:00 256.5 F 18.7
18:00 268.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.4
1:00 272.0 F 28.5
2:00 272.3 F 28.6
3:00 272.7 F 28.9
4:00 271.9 F 28.7

Min 271.6 F 28.2
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.3 F 0.0
5:00 232.9 F 0.0
10:00 247.3 F 0.0
15:00 237.9 F 0.0
20:00 228.7 F 0.0
25:00 229.9 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:60
HEATING: 18:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:12
CYCLE COMPLETE

05/24/2023 12:01 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____