
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.0uS / 1.2ppm
CYCLE NUMBER 001324
14:52 27/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:42
PRESSURIZING 3:44
STERILIZING 4:54
136.0 C 220kPa 4:54
136.3 C 224kPa 5:25
136.4 C 225kPa 5:55
136.4 C 225kPa 6:25
136.1 C 222kPa 6:55
136.1 C 223kPa 7:25
136.0 C 223kPa 7:55
136.3 C 224kPa 8:25
136.3 C 224kPa 8:25
VENTING 8:25
AIR DRYING 9:20
CYCLE COMPLETE 17:15

Digital Signature #
E21483CD42651E87
