
Midmark M11 - v1.0.5
Total Cycles: 2076

Name: RITTER2
Sterilizer ID: V2313228
02/22/2023 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.4 F 0.0
2:00 191.6 F 3.1
4:00 207.1 F 8.5
6:00 225.7 F 9.3
8:00 234.8 F 11.1
10:00 236.0 F 9.5
12:00 237.5 F 9.1
14:00 251.2 F 16.2
16:00 263.8 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.3 F 29.2
2:00 271.6 F 29.0
3:00 272.5 F 29.3
4:00 271.4 F 28.6

Min 271.3 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 229.8 F 0.0
10:00 215.7 F 0.0
15:00 216.4 F 0.0
20:00 224.4 F 0.0
25:00 238.7 F 0.0
30:00 228.3 F 0.0

Drying Complete

FILLING: 2:15
HEATING: 17:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

02/22/2023 12:52 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____