
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002329
9:45 31/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:18
PRESSURIZING 3:19
STERILIZING 4:26
136.0 C 220kPa 4:26
135.8 C 221kPa 4:56
136.4 C 223kPa 5:26
136.0 C 220kPa 5:56
136.4 C 224kPa 6:26
136.4 C 224kPa 6:56
136.4 C 224kPa 7:26
136.2 C 222kPa 7:56
136.2 C 222kPa 7:56
VENTING 7:56
AIR DRYING 8:51
CYCLE COMPLETE 9:16

Digital Signature #
5016B3709B120329
