
Midmark M11 - v1.0.0
Total Cycles: 2851

Name: RITTER 1
Sterilizer ID: V2178198
02/02/2023 03:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.8 F 0.0
2:00 192.6 F 4.8
4:00 207.1 F 9.8
6:00 222.6 F 6.0
8:00 231.1 F 10.0
10:00 231.2 F 7.8
12:00 242.6 F 13.2
14:00 242.6 F 11.9
16:00 255.7 F 18.8
18:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.4 F 29.4
2:00 271.8 F 28.9
3:00 271.2 F 28.7
4:00 272.4 F 29.5

Min 271.1 F 28.5
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 219.1 F 0.0
10:00 239.9 F 0.0
15:00 238.2 F 0.0
20:00 235.8 F 0.0
25:00 224.2 F 0.0
30:00 221.4 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 19:23
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

02/02/2023 04:19 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____