

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1801

Name: RITTER2  
Sterilizer ID: V2313228  
12/05/2022 12:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 151.5 F 0.0  
2:00 151.2 F 2.2  
4:00 166.6 F 5.5  
6:00 181.6 F 10.0  
8:00 215.8 F 5.8  
10:00 225.4 F 9.5  
12:00 230.9 F 8.6  
14:00 239.6 F 12.2  
16:00 239.1 F 10.7  
18:00 252.4 F 17.1  
20:00 263.0 F 23.6  
22:00 272.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.2  
1:00 272.7 F 30.1  
2:00 272.1 F 29.3  
3:00 271.5 F 29.4  
4:00 272.7 F 30.0

Min 271.3 F 29.0  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 233.1 F 0.0  
10:00 221.2 F 0.0  
15:00 230.7 F 0.0  
20:00 229.5 F 0.0  
25:00 227.5 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 22:15  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:00:19  
CYCLE COMPLETE

12/05/2022 01:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_