
Midmark M11 - v1.0.0
Total Cycles: 4621

Name: RITTER 1
Sterilizer ID: V2178198
06/10/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.2 F 0.0
2:00 141.9 F 1.1
4:00 158.7 F 3.5
6:00 161.1 F 7.0
8:00 182.2 F 11.0
10:00 208.9 F 3.9
12:00 215.3 F 7.7
14:00 227.8 F 12.1
16:00 231.2 F 8.1
18:00 243.2 F 13.5
20:00 241.9 F 11.7
22:00 253.9 F 17.8
24:00 264.7 F 24.1
26:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 272.5 F 29.4
2:00 271.5 F 28.6
3:00 272.4 F 29.5
4:00 272.2 F 29.1

Min 271.3 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 202.5 F 0.0
10:00 206.1 F 0.0
15:00 213.1 F 0.0
20:00 219.1 F 0.0
25:00 224.2 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:01
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:04:31
CYCLE COMPLETE

06/10/2024 11:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____