
Midmark M11 - v1.0.5
Total Cycles: 3358

Name: RITTER2
Sterilizer ID: V2313228
02/06/2024 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 83.0 F 0.0
4:00 91.2 F 0.0
6:00 115.5 F 1.0
8:00 141.1 F 3.6
10:00 168.9 F 6.9
12:00 205.5 F 10.6
14:00 219.1 F 5.8
16:00 224.0 F 7.0
18:00 232.7 F 10.8
20:00 231.4 F 7.4
22:00 241.0 F 11.4
24:00 236.8 F 9.1
26:00 249.8 F 15.2
28:00 260.3 F 21.3
30:00 270.0 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.3 F 29.1
2:00 271.4 F 28.9
3:00 272.5 F 29.0
4:00 271.2 F 28.9

Min 271.0 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 229.2 F 0.0
10:00 221.5 F 0.0
15:00 210.7 F 0.0
20:00 214.3 F 0.0
25:00 221.7 F 0.0
30:00 219.2 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:08:57
CYCLE COMPLETE

02/06/2024 10:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____