
Midmark M11 - v1.0.0
Total Cycles: 4880

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2024 01:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.9 F 0.0
2:00 167.6 F 0.4
4:00 182.7 F 2.6
6:00 206.4 F 7.2
8:00 220.1 F 12.3
10:00 226.7 F 7.5
12:00 239.6 F 13.2
14:00 239.6 F 11.0
16:00 237.8 F 9.6
18:00 252.3 F 16.6
20:00 264.3 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.7 F 29.1
2:00 271.8 F 28.4
3:00 272.1 F 29.0
4:00 272.5 F 28.9

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 210.1 F 0.0
10:00 224.1 F 0.0
15:00 223.0 F 0.0
20:00 228.9 F 0.0
25:00 230.8 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

09/09/2024 02:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____