
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002342
9:49 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:56
STERILIZING 4:04
136.5 C 226kPa 4:04
136.3 C 225kPa 4:34
136.4 C 224kPa 5:04
136.4 C 222kPa 5:34
136.4 C 223kPa 6:04
136.5 C 222kPa 6:34
136.2 C 224kPa 7:04
136.4 C 226kPa 7:34
136.4 C 226kPa 7:34
VENTING 7:34
AIR DRYING 8:26
CYCLE COMPLETE 8:58

Digital Signature #
A7AAB2E3F6A17252
