

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2124

Name: RITTER2  
Sterilizer ID: V2313228  
03/08/2023 09:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.9 F 0.0  
2:00 145.8 F 0.3  
4:00 147.7 F 3.0  
6:00 183.5 F 7.0  
8:00 202.7 F 12.0  
10:00 218.6 F 6.5  
12:00 231.9 F 11.5  
14:00 232.2 F 8.3  
16:00 241.2 F 11.3  
18:00 244.7 F 13.0  
20:00 257.0 F 19.7  
22:00 268.2 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.6 F 29.7  
2:00 271.7 F 28.9  
3:00 272.2 F 29.7  
4:00 272.4 F 29.5

Min 271.1 F 28.8  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 220.1 F 0.0  
10:00 229.7 F 0.0  
15:00 227.7 F 0.0  
20:00 224.2 F 0.0  
25:00 222.8 F 0.0  
30:00 221.5 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 23:16  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:01:19  
CYCLE COMPLETE

03/08/2023 10:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_