

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000292  
13:22 16/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:19  
STERILIZING 4:30  
136.0 C 228kPa 4:30  
136.3 C 230kPa 5:00  
136.4 C 233kPa 5:30  
136.4 C 233kPa 6:00  
136.4 C 232kPa 6:30  
136.1 C 229kPa 7:00  
136.3 C 232kPa 7:30  
136.1 C 231kPa 8:00  
136.1 C 231kPa 8:00  
VENTING 8:00  
AIR DRYING 8:52  
CYCLE COMPLETE 18:26  
-----

Digital Signature #  
C5FFA97C0CA2EDDD  
-----