
Midmark M11 - v1.0.5
Total Cycles: 1773

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.0 F 0.0
2:00 77.5 F 0.0
4:00 82.8 F 0.0
6:00 95.6 F 0.0
8:00 127.6 F 0.1
10:00 179.7 F 0.3
12:00 203.5 F 2.9
14:00 216.6 F 6.3
16:00 227.0 F 9.9
18:00 236.0 F 10.6
20:00 237.3 F 11.0
22:00 235.3 F 8.8
24:00 240.3 F 10.3
26:00 247.7 F 14.4
28:00 259.6 F 21.0
30:00 270.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.2 F 29.5
2:00 271.3 F 28.8
3:00 272.2 F 29.6
4:00 271.6 F 28.9

Min 271.3 F 28.7
Max 272.3 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.1 F 0.0
5:00 183.3 F 0.0
10:00 200.0 F 0.0
15:00 203.2 F 0.0
20:00 201.0 F 0.0
25:00 200.1 F 0.0
30:00 201.1 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:56
CYCLE COMPLETE

11/28/2022 08:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____