

-----  
Midmark M11 - v1.0.5  
Total Cycles: 901

Name: RITTER2  
Sterilizer ID: V2313228  
03/24/2022 10:51 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
2:00 184.9 F 2.2  
4:00 202.7 F 7.9  
6:00 223.9 F 12.4  
8:00 232.3 F 9.9  
10:00 234.5 F 8.6  
12:00 240.9 F 10.7  
14:00 248.9 F 15.0  
16:00 262.0 F 22.5  
18:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.8  
1:00 272.7 F 29.4  
2:00 271.2 F 28.7  
3:00 272.8 F 29.5  
4:00 271.9 F 28.9  
  
Min 271.1 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 236.1 F 0.0  
10:00 224.8 F 0.0  
15:00 229.4 F 0.0  
20:00 225.9 F 0.0  
25:00 227.2 F 0.0  
30:00 229.4 F 0.0  
  
Drying Complete

FILLING: 1:45  
HEATING: 18:13  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:07  
CYCLE COMPLETE

03/24/2022 11:48 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02