
Midmark M11 - v1.0.0
Total Cycles: 4387

Name: RITTER 1
Sterilizer ID: V2178198
03/20/2024 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.3 F 0.0
2:00 136.9 F 1.3
4:00 143.5 F 4.3
6:00 170.1 F 8.1
8:00 192.6 F 12.5
10:00 209.9 F 4.4
12:00 222.2 F 8.5
14:00 234.6 F 13.1
16:00 235.3 F 9.5
18:00 238.1 F 9.6
20:00 246.1 F 13.5
22:00 258.0 F 20.0
24:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.1
2:00 271.6 F 28.7
3:00 272.3 F 29.0
4:00 271.5 F 28.6

Min 271.4 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 225.6 F 0.0
10:00 231.0 F 0.0
15:00 227.7 F 0.0
20:00 228.3 F 0.0
25:00 226.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:05
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:03:34
CYCLE COMPLETE

03/20/2024 03:31 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____