
Midmark M11 - v1.0.0
Total Cycles: 4300

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2024 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.8 F 0.0
2:00 72.9 F 0.2
4:00 102.0 F 1.1
6:00 115.7 F 3.4
8:00 141.6 F 6.3
10:00 169.2 F 9.7
12:00 190.9 F 13.4
14:00 206.8 F 3.8
16:00 216.5 F 7.0
18:00 226.3 F 10.7
20:00 229.5 F 7.5
22:00 237.6 F 10.8
24:00 234.7 F 7.8
26:00 248.0 F 14.4
28:00 259.4 F 20.6
30:00 269.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 28.9
2:00 271.9 F 28.8
3:00 272.3 F 28.8
4:00 272.0 F 28.9

Min 271.5 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 221.8 F 0.0
10:00 229.0 F 0.0
15:00 230.5 F 0.0
20:00 223.5 F 0.0
25:00 221.8 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:59
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:09:16
CYCLE COMPLETE

02/14/2024 10:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____