
Midmark M11 - v1.0.0
Total Cycles: 1271

Name: RITTER 1
Sterilizer ID: V2178198
10/18/2021 06:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 71.8 F 0.0
4:00 95.9 F 0.0
6:00 102.5 F 0.1
8:00 128.2 F 0.1
10:00 163.5 F 0.2
12:00 186.8 F 0.6
14:00 204.1 F 3.7
16:00 214.3 F 6.6
18:00 224.0 F 9.8
20:00 234.0 F 13.5
22:00 234.5 F 9.4
24:00 241.0 F 10.9
26:00 243.7 F 12.3
28:00 241.1 F 10.8
30:00 253.7 F 17.1
32:00 264.3 F 23.3
34:00 271.8 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.8
2:00 271.5 F 28.0
3:00 272.5 F 28.9
4:00 272.1 F 28.5

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 200.6 F 0.0
10:00 195.7 F 0.0
15:00 197.5 F 0.0
20:00 187.4 F 0.0
25:00 185.3 F 0.0
30:00 185.5 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 34:07
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:12:15
CYCLE COMPLETE

10/18/2021 07:58 AM
ACCEPTED: X
REJECTED:
Operator: U01