
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.5uS / 1.6ppm
CYCLE NUMBER 003681
10:40 25/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:21
136.0 C 220kPa 3:21
136.1 C 224kPa 3:51
136.4 C 224kPa 4:21
136.4 C 223kPa 4:51
136.3 C 225kPa 5:21
136.4 C 224kPa 5:51
136.1 C 219kPa 6:21
136.4 C 224kPa 6:51
136.4 C 224kPa 6:51
VENTING 6:51
AIR DRYING 7:37
CYCLE COMPLETE 9:00

Digital Signature #
9A2BD1D0D03F60B8
