

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1935

Name: RITTER 1  
Sterilizer ID: V2178198  
05/11/2022 04:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
2:00 182.7 F 3.7  
4:00 198.2 F 8.4  
6:00 215.3 F 10.0  
8:00 229.8 F 9.3  
10:00 234.6 F 9.0  
12:00 244.0 F 12.6  
14:00 242.4 F 11.3  
16:00 255.6 F 18.1  
18:00 266.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.4 F 28.6  
2:00 271.7 F 28.4  
3:00 272.3 F 28.6  
4:00 271.6 F 28.2  
  
Min 271.4 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.8 F 0.0  
5:00 234.1 F 0.0  
10:00 231.4 F 0.0  
15:00 231.3 F 0.0  
20:00 231.3 F 0.0  
25:00 228.8 F 0.0  
30:00 226.3 F 0.0  
  
Drying Complete

FILLING: 1:60  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:37  
CYCLE COMPLETE

05/12/2022 07:06 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02