

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5072

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2024 10:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 71.8 F 0.0  
4:00 95.4 F 0.0  
6:00 145.8 F 0.1  
8:00 158.7 F 0.2  
10:00 172.2 F 0.7  
12:00 189.4 F 3.7  
14:00 204.5 F 6.4  
16:00 213.8 F 9.7  
18:00 223.6 F 13.3  
20:00 226.6 F 7.5  
22:00 237.5 F 12.1  
24:00 236.2 F 9.1  
26:00 240.0 F 10.2  
28:00 245.9 F 13.1  
30:00 257.0 F 18.9  
32:00 267.1 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.5 F 28.9  
2:00 271.5 F 28.1  
3:00 272.4 F 28.9  
4:00 272.0 F 28.5  
  
Min 271.4 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.1 F 0.0  
5:00 188.3 F 0.0  
10:00 185.5 F 0.0  
15:00 187.7 F 0.0  
20:00 189.3 F 0.0  
25:00 214.5 F 0.0  
30:00 222.0 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 33:35  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:11:55  
CYCLE COMPLETE

11/08/2024 11:18 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_