
Midmark M11 - v1.0.0
Total Cycles: 2726

Name: RITTER 1
Sterilizer ID: V2178198
12/28/2022 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.2 F 0.0
2:00 126.6 F 0.1
4:00 158.9 F 1.9
6:00 183.6 F 5.8
8:00 206.3 F 11.5
10:00 219.0 F 5.6
12:00 232.5 F 10.5
14:00 232.1 F 7.2
16:00 245.0 F 13.2
18:00 243.4 F 11.8
20:00 256.9 F 18.8
22:00 268.9 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 272.7 F 28.7
2:00 271.5 F 28.0
3:00 272.5 F 28.4
4:00 271.7 F 28.2

Min 271.3 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.7 F 0.0
5:00 203.3 F 0.0
10:00 219.0 F 0.0
15:00 222.6 F 0.0
20:00 222.1 F 0.0
25:00 221.7 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:53
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

12/28/2022 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____