
Midmark M11 - v1.0.0
Total Cycles: 1827

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.8 F 0.0
2:00 182.2 F 4.2
4:00 195.7 F 9.3
6:00 221.6 F 6.8
8:00 232.2 F 9.5
10:00 234.3 F 8.7
12:00 243.8 F 12.7
14:00 242.5 F 11.6
16:00 256.1 F 18.7
18:00 268.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.1
2:00 271.4 F 28.3
3:00 272.4 F 29.2
4:00 272.2 F 28.9

Min 271.2 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 201.0 F 0.0
10:00 208.4 F 0.0
15:00 215.1 F 0.0
20:00 218.8 F 0.0
25:00 218.0 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:37
CYCLE COMPLETE

04/14/2022 02:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____