

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1622

Name: RITTER2  
Sterilizer ID: V2313228  
10/19/2022 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
2:00 195.1 F 3.5  
4:00 194.8 F 8.4  
6:00 217.3 F 10.8  
8:00 228.6 F 9.6  
10:00 234.4 F 8.9  
12:00 244.0 F 13.2  
14:00 243.0 F 12.4  
16:00 255.9 F 19.2  
18:00 267.3 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.3  
1:00 272.3 F 30.1  
2:00 272.1 F 29.6  
3:00 271.1 F 29.2  
4:00 271.9 F 30.1  
  
Min 271.1 F 29.1  
Max 272.4 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 214.1 F 0.0  
10:00 228.1 F 0.0  
15:00 231.8 F 0.0  
20:00 231.8 F 0.0  
25:00 228.2 F 0.0  
30:00 227.2 F 0.0  
  
Drying Complete

FILLING: 1:52  
HEATING: 19:29  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:35  
CYCLE COMPLETE

10/19/2022 10:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_