
Midmark M11 - v1.0.0
Total Cycles: 5234

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2025 11:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.2 F 0.0
2:00 159.9 F 2.6
4:00 178.6 F 6.6
6:00 205.2 F 11.4
8:00 219.8 F 5.4
10:00 232.1 F 10.5
12:00 231.5 F 7.1
14:00 244.0 F 12.9
16:00 241.4 F 11.1
18:00 254.2 F 17.5
20:00 265.0 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 271.7 F 28.4
2:00 272.2 F 28.7
3:00 272.1 F 29.0
4:00 271.1 F 28.3

Min 271.1 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 229.3 F 0.0
10:00 228.6 F 0.0
15:00 225.4 F 0.0
20:00 226.8 F 0.0
25:00 236.4 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:46
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:00:29
CYCLE COMPLETE

01/10/2025 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____